



Original article

In vivo evaluation of porous hydrogel pins to fill osteochondral defects in rabbits[☆]

Túlio Pereira Cardoso ^{a,*}, André Petry Sandoval Ursolino ^a, Pamela de Melo Casagrande ^a, Edie Benedito Caetano ^a, Daniel Vinicius Mistura ^b, Eliana Aparecida de Rezende Duek ^a

^a Faculdade de Ciências Médicas e da Saúde de Sorocaba, Pontifícia Universidade Católica de São Paulo, Sorocaba, SP, Brazil

^b Universidade Federal de São Carlos, Sorocaba, SP, Brazil

ARTICLE INFO

Article history:

Received 14 January 2016

Accepted 29 March 2016

Available online xxx

Keywords:

Articular cartilage

Hydrogels/chemistry

Rabbits

ABSTRACT

Objective: This experimental study aimed to evaluate the biological performance of poly (L-co-D, L-lactic acid)-co-trimethylene carbonate/poly (vinyl alcohol) (PLDLA-co TMC/PVA), hydrogel scaffolds, as an implant in the filling (and not in the repair) of osteochondral defects in New Zealand rabbits, assessing the influence of the material in tissue protection in vivo.

Methods: Twelve rabbits were divided into groups of nine and 16 weeks. In each animal, an osteochondral defect was created in both medial femoral condyles. In one knee, a hydrogel scaffold was implanted (pin group) and in the other, the defect was maintained (control group). A histological analysis of the material was performed after euthanasia.

Results: The condyles of the pin group showed no inflammatory reaction and were surrounded by a fibrous capsule. The control group presented higher bone growth in the areas of the defect, but with disorganized articular cartilage, evident fibrosis, bone exposure, atrophy, and proliferation of synovial membrane.

Conclusion: The hydrogel pins are promising in filling osteochondral defects, generally do not cause inflammatory reactions, and are not effective in the repair of osteochondral defects.

© 2016 Sociedade Brasileira de Ortopedia e Traumatologia. Published by Elsevier Editora Ltda. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

[☆] Study conducted at the Biomaterials Laboratory, Faculty of Medical and Health Sciences in Sorocaba, Pontifícia Universidade Católica de São Paulo, Sorocaba, SP, Brazil.

* Corresponding author.

E-mail: tuliopecardoso@gmail.com (T.P. Cardoso).

<http://dx.doi.org/10.1016/j.rboe.2016.03.009>

2255-4971/© 2016 Sociedade Brasileira de Ortopedia e Traumatologia. Published by Elsevier Editora Ltda. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

Avaliação do desempenho *in vivo* de pinos porosos de hidrogel para preenchimento de defeito osteocondral em coelhos

RESUMO

Palavras-chave:

Cartilagem articular
Hidrogéis/química
Coelhos

Objetivo: Trabalho experimental para avaliar o desempenho biológico de arcabouços de hidrogel poli (L-co-D, L ácido lático)-co-trimetileno carbonato/poli (álcool vinílico) (PLDLA-co-TMC/PVA) como implante no preenchimento, e não no reparo, de defeito osteocondral em coelhos Nova Zelândia e verificar a influência do material na proteção tecidual *in vivo*.

Métodos: Foram usados 12 coelhos divididos em grupos de nove e 16 semanas. Em cada animal foi criado um defeito osteocondral em ambos os côndilos femorais mediais, em um foi implantado um arcabouço de hidrogel (grupo pino) e no outro foi mantido o defeito (grupo controle). Após o sacrifício dos animais, foi feita análise histológica do material.

Resultados: Os côndilos do grupo pino não evidenciaram reação inflamatória e estavam rodeados por cápsula fibrosa. Já no grupo controle, uma maior proliferação óssea foi observada nas áreas do defeito, porém com cartilagem articular desorganizada, fibrose evidente, atrofia com exposição óssea e proliferação de membrana sinovial.

Conclusão: Os pinos de hidrogel são promissores na função de preenchimento de defeitos osteocondrais, não ocasionam, de modo geral, reação inflamatória e não são eficazes no reparo de defeitos osteocondrais.

© 2016 Sociedade Brasileira de Ortopedia e Traumatologia. Publicado por Elsevier Editora Ltda. Este é um artigo Open Access sob uma licença CC BY-NC-ND (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

Introduction

Osteoarthritis is one of the diseases that most commonly affect humans.¹ Its prevalence increases with age, being common after 60 years.² The articular cartilage and the subchondral bone form a lubrication, stabilization, and uniform load distribution system, absorbing shocks and allowing movement with low friction for several decades.²⁻⁴ Thus, cartilage protects subchondral bone from high stress and reduces normal contact pressure.^{2,5} Degraded cartilage evolves to joint pain, stiffness, and decreased movement. Due to low chondral regeneration capacity, osteoarthritis is one of the most important problems in orthopedics. With increasing human longevity and the practice of sports in recent decades, osteochondral injuries have been increasingly observed.²

Normal joint cartilage features a solid phase consisting mostly of collagen and proteoglycans (15–32%) and a fluid phase composed predominantly of water (68–85%).² Hyaline cartilage consists of 10% cells (chondrocytes) and a dense extracellular matrix, composed of 60–80% water, 10–20% type II collagen fibers, and 10–15% proteoglycans. The mechanical properties of the articular cartilage allow it to transmit loads of approximately eight times the body weight, due to exudation and movement of the fluid through the pores of cartilage, conferring a friction coefficient of 0.008 μ (mi).^{4,6}

The subchondral bone is a thin layer of dense, hard bone in contact with the articular cartilage, while trabecular bone is composed of an abundant matrix (collagen fibers and minerals) that serves to transmit loads.⁶

Being avascular, the cartilage depends on the vascularization of the bone marrow for the migration of mesenchymal

cells responsible for the healing process.^{7,8} Furthermore, superficial lesions of the articular cartilage without subchondral bone involvement have little intrinsic repair capacity.⁹

Both in primary and secondary osteoarthritis, cartilage is the tissue that undergoes the greatest damage. Morphological changes to the cartilage include loss of its homogeneous nature, fragmentation, fibrillation, fissures, and ulcerations. With disease progression, occasionally no cartilage remains and areas of the subchondral bone become exposed.¹⁰

Three stages can be considered in the process of tissue regeneration: necrosis, inflammation, and repair.^{11,12} Nonetheless, healing of lesions restricted to the hyaline cartilage may not occur this way.¹³ Superficial lesions of the articular cartilage that do not reach the subchondral bone tend not to heal.⁹ In these lesions, there is a degeneration of the cartilage from the surface area; thin portions of collagen fibers with scaly appearance are observed. With lesion progression, vertical cracks with an uneven and dull appearance can be observed in the articular cartilage.^{4,14}

Chondral lesions trigger an inflammatory process and the hematoma quickly organizes itself into fibrin clots, white blood cells, and bone marrow elements.

Undifferentiated bone marrow and vascular endothelium cells are converted into primitive fibroblasts, which, with input of capillaries and fibrin clots, turn into vascular fibroblastic repair tissue.¹¹ Depending on the mechanical and biological stimuli, this fibrocartilage tissue will form a cartilaginous tissue.¹³ Newly formed bone migrates from the base of the defect to the articular surface in the area in contact with subchondral bone. Fibrocartilaginous tissue fills the transition zone and interrupts the formation of bone tissue.^{4,11,14}

دريافت فوري

متن كامل مقاله



- ✓ امكان دانلود نسخه تمام مقالات انگلیسي
- ✓ امكان دانلود نسخه ترجمه شده مقالات
- ✓ پذيرش سفارش ترجمه تخصصي
- ✓ امكان جستجو در آرشيو جامعى از صدها موضوع و هزاران مقاله
- ✓ امكان دانلود رايگان ۲ صفحه اول هر مقاله
- ✓ امكان پرداخت اينترنتى با کليه کارت های عضو شتاب
- ✓ دانلود فوري مقاله پس از پرداخت آنلاين
- ✓ پشتيباني كامل خريد با بهره مندي از سيسitem هوشمند رهگيری سفارشات